

September - 2020

classmate

Date _____

Page _____

Date	Department	Complaints	Rectified date	Sign
1/9/20	Orthodontics	Chair bulb change	3/9/20	Epl
3/9/20	Periodontics	Three way Syringe	5/9/20	Epl
3/9/20	Pedodontics	Light Switches	4/9/20	Epl
7/9/20	Oral and maxillofacial surgery	Airtube	8/9/20	Dr
8/9/20	Public health dentistry	Foot control	10/9/20	Ch
8/9/20	Oral pathology	Light Switches	11/9/20	Epl
10/9/20	OMR	Chair bulb change	11/9/20	Ch
11/9/20	Cons and Endo	Aerator coupling	12/9/20	Epl
14/9/20	Prosthodontics	Booster ring	12/9/20	Epl
14/9/20	Orthodontics	Water retract valve	15/9/20	Ch
15/9/20	PHD	Three way Syringe	17/9/20	f
18/9/20	Pedodontics	Airtube	18/9/20	Epl
18/9/20	OMFS	Chair bulb change	20/9/20	Epl
19/9/20	OP	Light Switches	21/9/20	f
20/9/20	Prosthodontics	Booster ring	21/9/20	Ch
22/9/20	Cons and Endo	Aerator coupling	22/9/20	Ch
24/9/20	OMFS	Light Switches	25/9/20	Epl
25/9/20	Orthodontics	Foot control	26/9/20	Epl
26/9/20	Pedodontics	Three way Syringe	27/9/20	f
27/9/20	Periodontics	Water retract valve	27/9/20	Epl
28/9/20	OMFS	Light Switches	29/9/20	f
29/9/20	Oral pathology	Chair bulb change	30/9/20	f
31/9/20	Prosthodontics	Foot control	31/9/20	Epl
3/9/20	PHD	Airtube	31/9/20	f

October 2020

classmate

Date _____

Page _____

Date	Department	Complaints	Rectified etc	Signature
1/10/20	Orthodontics	Chair bulb damage	2/10/20	[Signature]
1/10/20	OMFS	Water retractor valve	3/10/20	[Signature]
2/10/20	Cores and Endo	Three way syringe	2/10/20	[Signature]
3/10/20	Periodontics	Light switches	5/10/20	[Signature]
3/10/20	Orthodontics	Booster o ring	3/10/20	[Signature]
3/10/20	Pedodontics	Air tube	5/10/20	[Signature]
5/10/20	Pedodontics	Aerator coupling	6/10/20	[Signature]
6/10/20	Periodontics	Light switches	8/10/20	[Signature]
8/10/20	OMFS	Chair bulb valve	9/10/20	[Signature]
8/10/20	Periodontics	Air tube	10/10/20	[Signature]
7/10/20	Cores and Endo	Water retractor valve	11/10/20	[Signature]
12/10/20	Oral Medicine	Foot control	13/10/20	[Signature]
14/10/20	Orthodontics	Foot control	17/10/20	[Signature]
17/10/20	Oral pathology	Light switches	17/10/20	[Signature]
18/10/20	PHD	Air tube	19/10/20	[Signature]
20/10/20	OMFS	Booster o ring	20/10/20	[Signature]
20/10/20	OMR	Aerator coupling	20/10/20	[Signature]
21/10/20	Cores and Endo	Water retractor valve	23/10/20	[Signature]
24/10/20	Pedodontics	Foot control	25/10/20	[Signature]
25/10/20	Pedodontics	Air tube	25/10/20	[Signature]
25/10/20	PHD	Three way syringe	26/10/20	[Signature]
26/10/20	Periodontics	Foot control	27/10/20	[Signature]
27/10/20	OMFS	Booster o ring	28/10/20	[Signature]
28/10/20	Orthodontics	Light switches	30/10/20	[Signature]

November - 2020.

Date _____

Page _____

Date	Department	Complaints	Rectified date	Sign
2/11/20	Orthodontics	Water retriever valve	2/11/20	Yip
2/11/20	Periodontics	Three way syringe	3/11/20	Yip
5/11/20	Pedodontics	Light Switches	6/11/20	Yip
6/11/20	Periodontics	Three way syringes	9/11/20	Yip
6/11/20	Oral pathology	Chair bulb damage	9/11/20	Yip
7/11/20	Cos and Endo	Booster wiring	9/11/20	Yip
7/11/20	OMPS	Booster wiring	9/11/20	Yip
9/11/20	Orthodontics	Acrotor coupling	10/11/20	Yip
9/11/20	OMPS	Water retriever valve	20/11/20	Yip
10/11/20	Cos and Endo	Acrotor coupling	12/11/20	Yip
11/11/20	OMR	Foot control	12/11/20	Yip
12/11/20	PHD	Three way syringe	13/11/20	Yip
13/11/20	Oral pathology	Air subc	24/11/20	Yip
15/11/20	OMPS	Light switches	14/11/20	Yip
14/11/20	OMR	Foot control	17/11/20	Yip
16/11/20	Oral pathology	Booster wiring	16/11/20	Yip
16/11/20	Pedodontics	Light Switches	19/11/20	Yip
17/11/20	Orthodontics	Foot control	19/11/20	Yip
17/11/20	PHD	Acrotor coupling	18/11/20	Yip
20/11/20	Pedodontics	Light Switches	21/11/20	Yip
23/11/20	OMR	Chair bulb damage	24/11/20	Yip
24/11/20	Periodontics	Water retriever	25/11/20	Yip
26/11/20	OMPS	Light Switches	30/11/20	Yip
27/11/20	Orthodontics	Light Switches	30/11/20	Yip

December 2020

classmate

Date _____

Page _____

Date	Department	Complaints	Rectified date	Signature
1/12/20	Orthodontics	Chair bulb damage	2/12/20	[Signature]
1/12/20	Pedodontics	Three way Syng	2/12/20	[Signature]
2/12/20	Pedodontics	Aerator coupling	3/12/20	[Signature]
8/12/20	Periodontics	Chair bulb damage	4/12/20	[Signature]
3/12/20	Periodontics	Air tube	4/12/20	[Signature]
7/12/20	Oral Pathology	Aerator coupling	8/12/20	[Signature]
8/12/20	Prosthodontics	Chair bulb damage	10/12/20	[Signature]
9/12/20	Orthodontics	Air tube	12/12/20	[Signature]
12/19/20	OMFS	Foot control	15/12/20	[Signature]
13/12/20	PTP	Foot control	15/12/20	[Signature]
15/12/20	OMR	Light switches	16/12/20	[Signature]
18/12/20	Oral Pathology	Aerator coupling	19/12/20	[Signature]
17/12/20	OMR	Foot control	17/12/20	[Signature]
18/12/20	OMFS	Booker's ring	19/12/20	[Signature]
21/12/20	Pedodontics	light switches	22/12/20	[Signature]
21/12/20	Periodontics	light switch	21/12/20	[Signature]
22/12/20	Prosthodontics	Aerator coupling	24/12/20	[Signature]
23/12/20	OMFS	Chair bulb damage	24/12/20	[Signature]
24/12/20	Pedodontics	Air tube	26/12/20	[Signature]
25/12/20	Orthodontics	Foot control	26/12/20	[Signature]
28/12/20	PTP	Three way Syringe	29/12/20	[Signature]
28/12/20	Orthodontics	Water control Valve	29/12/20	[Signature]
29/12/20	OMFS	light switches	30/12/20	[Signature]
30/12/20	OMR	Air tube	31/12/20	[Signature]

January 2021

classmate

Date _____

Page _____

Date	Department	Complaints	Resolved date	Sign
1/1/21	Orthodontic	Three way syringe	1/1/21	Sh
1/1/21	OMFS	light Switches	2/1/21	Gya
1/1/21	PHD	Aerator coupling	2/1/21	Gya
2/1/21	OMR	Chair bulbs dange	4/1/21	Gya
4/1/21	OMFS	Aerator coupling	5/1/21	Gya
5/1/21	Pedodontics	Air tube	6/1/21	Gya
6/1/21	Periodontics	light Switches	8/1/21	Gya
7/1/21	OMR	Air tube	8/1/21	Gya
8/1/21	Pedodontics	Aerator coupling	9/1/21	Gya
11/1/21	Periodontics	Air tube	13/1/21	Gya
11/1/21	End. pathology	Foot control	13/1/21	Gya
13/1/21	Prosthodontics	light Switches	15/1/21	Gya
15/1/21	PHD	Three way syringe	18/1/21	Gya
18/1/21	OMR	Water retainer valve	19/1/21	Gya
18/1/21	Prosthodontics	Booster lining	18/1/21	Gya
21/1/21	OMFS	Water retainer valve	22/1/21	Gya
22/1/21	Orthodontics	light Switches	23/1/21	Gya
23/1/21	Periodontics	Air tube	25/1/21	Gya
25/1/21	Prosthodontics	Chair bulbs dange	26/1/21	Gya
26/1/21	Pedodontics	light Switches	27/1/21	Gya
28/1/21	OMR	Air tube	29/1/21	Gya
30/1/21	Prosthodontics	Foot control	2/2/21	Gya
30/1/21	PHD	Aerator coupling	1/2/21	Gya
30/1/21	OMFS	OMFS	30/1/21	Gya

February 2021

classmate

Date _____

Page _____

Date	Department	Complaints	Detected Date	Sign
1/2/21	Orthodontics	Chair bulb damage	1/2/21	Qm
1/2/21	PHO	Thru by Syringe	2/2/21	Qm
2/2/21	Prosthodontics	Aerator coupling	2/2/21	Qm
2/2/21	Periodontics	Chair bulb damage	3/2/21	Qm
3/2/21	Pedodontics	Air tube	4/2/21	Qm
4/2/21	OMFs	Booster or syringe	6/2/21	A
5/2/21	OMK	Light switches	5/2/21	H
5/2/21	Oral pathology	Aerator coupling	8/2/21	H
5/2/21	Cons and Endo	Water extractor valve	3/2/21	H
8/2/21	Prosthodontics	Air tube	8/2/21	H
9/2/21	Cons and Endo	Water extractor valve	10/2/21	H
10/2/21	Orthodontics	Light switches	11/2/21	H
11/2/21	PHO	Booster or syringe	12/2/21	H
13/2/21	Periodontics	Light switches	15/2/21	Qm
16/2/21	Pedodontics	Aerator coupling	17/2/21	Qm
17/2/21	Prosthodontics	Air tube	17/2/21	Qm
18/2/21	PHO	Water extractor valve	19/2/21	H
20/2/21	OMFs	Aerator coupling	22/2/21	Qm
22/2/21	Periodontics	Air tube	23/2/21	H
23/2/21	Pedodontics	Light switches	24/2/21	H
25/2/21	OMR	Chair bulb damage	25/2/21	H
25/2/21	Prosthodontics	Light switches	26/2/21	Qm
26/2/21	Pedodontics	Water extractor valve	26/2/21	H
26/2/21	Periodontics	Air tube	27/2/21	H

March 2021

classmate

Date _____

Page _____

Date	Department	Complaints	Rectified date	Sign
1/3/21	Cons and Endo	light switches	1/3/21	Gpre
3/3/21	OMR	three way syringe	4/3/21	Gpre
5/3/21	Periodontics	Chair bulb damage	6/3/21	Gpre
6/3/21	Orthodontics	Booster o ring	6/3/21	Gpre
7/3/21	Pedodontics	Aerator coupling	8/3/21	Gpre
9/3/21	Periodontics	light switches	10/3/21	Gpre
10/3/21	Pedodontics	Aerator coupling	11/3/21	Gpre
12/3/21	OMR	Booster o ring	15/3/21	Gpre
13/3/21	OMFS	light switches	13/3/21	Gpre
14/3/21	Cons and Endo	Air tube	12/3/21	Gpre
15/3/21	PHD	Aerator coupling	17/3/21	Gpre
17/3/21	Oral pathology	three way syringe	20/3/21	Gpre
20/3/21	Oral pathology	light switches	22/3/21	Gpre
21/3/21	pedodontics	Air tube	23/3/21	Gpre
22/3/21	PHD	light switches	26/3/21	Gpre
24/3/21	prosthodontics	Water retactor valve	27/3/21	Gpre
26/3/21	OMR	three way syringe	27/3/21	Gpre
27/3/21	prosthodontics	Booster o ring	27/3/21	Gpre
29/3/21	periodontics	Chair bulb damage	29/3/21	Gpre
29/3/21	Cons and Endo	light switches	28/3/21	Gpre
28/3/21	OMFS	three way syringe	31/3/21	Gpre
29/3/21	PHD	Chair bulb damage	29/3/21	Gpre
29/3/21	OMR	Air tube	31/3/21	Gpre
30/3/21	Periodontics	Aerator coupling	29/3/21	Gpre

April - 2021

Date	Department	Complaints	Rectified date	Signature
1/9/21	Orthodontics	Chair Bulb damage	1/9/21	[Signature]
2/9/21	Prosthodontics	Booster orange	3/9/21	[Signature]
4/9/21	Periodontics	Foot control	3/9/21	[Signature]
2/9/21	Cons & Endo	light switches	6/9/21	[Signature]
5/9/21	Prosthodontics	3 way syringe	5/9/21	[Signature]
6/9/21	Periodontics	Air tube	7/9/21	[Signature]
7/9/21	Pedodontics	Foot control	8/9/21	[Signature]
7/9/21	OMR	light switches	9/9/21	[Signature]
10/9/21	Cons & Endo	Water retractor valve	12/9/21	[Signature]
13/9/21	Pedodontics	Foot control	14/9/21	[Signature]
13/9/21	OMR	Chair Bulb damage	13/9/21	[Signature]
16/9/21	OMFS	Chair Bulb damage	16/9/21	[Signature]
17/9/21	Cons & Endo	Water retractor	17/9/21	[Signature]
19/9/21	Pedodontics	3 way syringe	19/9/21	[Signature]
19/9/21	OMFS	Air tube	19/9/21	[Signature]
20/9/21	OMR	light switch	20/9/21	[Signature]
21/9/21	PHD	Air tube	21/9/21	[Signature]
21/9/21	OMR	Chair Bulb	21/9/21	[Signature]
22/9/21	Pedodontics	Chair Bulb damage	23/9/21	[Signature]
23/9/21	OMFS	light chair damage	24/9/21	[Signature]
23/9/21	Prosthodontics	Foot control	26/9/21	[Signature]
27/9/21	Pedodontics	light switches	28/9/21	[Signature]
29/9/21	Periodontics	Foot control	28/9/21	[Signature]
30/9/21	Orthodontics	light switches	30/9/21	[Signature]

August 2020

Date	Department	Complaints	Rectified date and Span.	Sign:
03/08/20	Oral Medicine and Radiology	3 - Fans Repair	05/08/20	Ramana
	Prosthodontics	Geyusers - 1	05/08/20	Prakash
	Microbiology	Switch board - 1	06/08/20	Ramana
	Periodontics	A-C Repair - 1	06/08/20	Verma
	Oral pathology	Lights repair - 2	05/08/20	Ramana
	Biochemistry	2 - Fans Repair	05/08/20	Ramana
12/08/20	Orthodontics	Lights repair - 3	14/08/20	Ramana
	conservative and endodontics	wiring	14/08/20	Prakash
	public health dentistry	Switch board	14/08/20	Ramana
	Periodontics	Projector	14/08/20	Praveen
	Pharmacology	Fan repair - 2	14/08/20	Prakash

Date	Department	Complaints	Rectified date and span	Sign
	Anatomy	lights repair-5	17/08/20	vinay
18/08/20	OMFS	lights repair-1	19/08/20	Ramana
	Physiology	Fans repair-2	19/08/20	Ramana
	Gen. Pathology	lights repair-1	19/08/20	Prakash
	Periodontics	Projector-1	20/08/20	Pranetha
	Orthodontics	Refrigerator-1	20/08/20	Quesh
24/08/20	Prosthodontics	Switch Board-1	25/08/20	Ramana
	Microbiology	Fans - 2	25/08/20	Prakash
	oral pathology	Fans- 1	26/08/20	Ramana
	Biochemistry	lights repair-3	26/08/20	Prakash
	Public health dentistry	wiring	26/08/20	Ramana

September

Date	Department	Complaints	Received date	Sign
16/9/20	OMFS	Fans repairs - 3.	21/9/20	Ramona
	Physiology	Tap repair	21/9/20	Appalas
	Pedodontics	wiring	21/9/20	Ramona
	Cons & Endo	Refrigerator	21/9/20	Queesh
	Prosthodontics	Leakage problem	31/9/20	Appalas
	Pedodontics	Pipe repair	21/9/20	Ram
09/10/20	OMR	Tap repair	10/10/20	Ram
	Anatomy	leakage Problem.	10/10/20	Appalas
	Gen. Pathology	Switch Board - 1	11/10/20	Ramona
	Orthodontics	wiring.	11/10/20	Queesh
	Pharmacology	light repair.	11/10/20	Ramona

Date	Department	Complaints	Received date	Sign.
14/09/20	Public health dentistry	lights repairs - 1	15/09/20	Ramana
	Pedodontics	Fans repairs - 2	15/09/20	Rumana
	Oral pathology	Tap repairs - 1	16/09/20	Aparna
	Microbiology	Refrigerator repairs - 1	16/09/20	Dush
	Prosthodontics	Pipe repairs - 2	16/09/20	Aparna
22/09/20	Cons and endodontics	Projector	23/09/20	Pranethi
	Pedodontics	AC repairs	23/09/20	Vandana
	Orthodontics	Projector	23/09/20	Pranethi
	Anatomy	lights repair - 3	24/09/20	Ramana
	OMR	Fan repairs - 1	24/09/20	Pranethi
	OMFS	Tap repairs - 2	24/09/20	Aparna

Date	Department	Complaints	Rectified date	San
28/09/20	Prosthodontics	Cigets Repair - 4	29/09/20	Ramkumar
	Biochemistry	pipe leakage	29/09/20	Rajao
	Periodontics	projector	29/09/20	Ramkumar
	plainsmology	fan Repair	30/09/20	Rajao
	public health dentistry	Switch Board - 1	30/09/20	Ramkumar
	Gen. pathology	Wiring	30/09/20	Rajao

October

Date	Department	Complaints	Received date	Sign
07/10/20	Cos and Endocrin	Projector	08/10/20	Praveen
	Pedodontics	AC Repair	09/10/20	Subbu
	Anatomy	Fan Repair	9/10/20	Ramesh
	Orthodontics	Ceaser Repair	08/10/20	Rishabh
	OMR	Fan Repair	9/10/20	Ramesh
	OMF	AC Repair	09/10/20	Subbu
12/10/20	Physiology	lights repair - 3	13/10/20	Rishabh
	Biochemistry	Tap Repair - 2	14/10/20	Appar
	DADH	lights repair - 2	13/10/20	Rishabh
	Dial Pathology	Projector	14/10/20	Praveen

Date	Department	Complaint	Received date	Signature
	Periodontics	Fan repair	14/10/20	Ramana
	Public Health Dentistry	Wiring	13/10/20	Rakesh
21/10/20	Periodontics	Pipe leakage	21/10/20	Appar
	OMFS	AC Repair	23/10/20	Subha
	Orthodontics	Refrigerator	23/10/20	Arush
	OMR	Fan Repair	22/10/20	Ramana
	Periodontics	Ceyses repair	21/10/20	Rakesh
	Microbiology	Refrigerator	22/10/20	Suresh
	Anatomy	Tap repair-2	22/10/20	Appar
27/10/20	Caus and Endodontics	Refrigerator	28/10/20	Suresh
	DADM	Fan Repair	28/10/20	Rama
	physiology	Tap Repair	29/10/20	Appar

Date	Department	Complaints	Accused date	By
29/10/20	Biochemistry	Cigarettes Repair - 3	29/10/20	Kamran
29/10/20	Prosthodontics	Geyser Repair	29/10/20	Prakash
28/10/20	Public Health Dentistry	Switchboard - 2	28/10/20	Ajith
28/10/20	Dental	Dental	28/10/20	
28/10/20	Dental	Dental	28/10/20	
28/10/20	Dental	Dental	28/10/20	
28/10/20	Dental	Dental	28/10/20	
28/10/20	Dental	Dental	28/10/20	
28/10/20	Dental	Dental	28/10/20	
28/10/20	Dental	Dental	28/10/20	
28/10/20	Dental	Dental	28/10/20	
28/10/20	Dental	Dental	28/10/20	

November

Date	Department	Complaint	Received date	By
03/11/20	Prostodontics	Jaw Repair - 2	04/11/20	Ranana
	Orthodontics	light repair - 2	04-11-20	Ranana
	Anatomy	Tap Repair - 2	05/11/20	Apparu
	OMR	pipe Repair	05/11/20	Apparu
	Cons & Endo	Projector	05/11/20	pranath
	OMFS	Switch Board - 2	05/11/20	Deekhat
11/11/20	OMR	Electrical wiring	12/11/20	Prakash
	Pedodontics	pipe Repair	12/11/20	Apparu
	Anatomy	fan Repair	13/11/20	Ranana
	periodontics	projector Repair	12/11/20	pranath
	Biochemistry	light Repair - 2	14/11/20	Ranana

DATE	Department	complaint	Received date	Signature
12/11/20	OMFS	faust Repair	18/11/20	Prakash
	Orthodontics	Refrigerator	19/11/20	Beech
	public Health dentistry	Ac Repair	19/11/20	Prash
	physiology	pipe Repair	18/11/20	Prash
	oral pathology	Tap Repair	18/11/20	Prash
	Anatomy	Switch Board	19/11/20	Prash
25/11/20	Biochemistry	pipe Repair	26/11/20	Prash
	prosthodontics	light Repair	26/11/20	Prash
	OMR	Switch Board	27/11/20	Prash
	Cons & Endo	Ac Repair	27/11/20	Venkat
	pedodontics	faust Repair	27/11/20	Prash

December

DATE	Department	work done	Rectified date	Sign
<u>1/12/20</u>	prosthodontics	pipe leakage	2/12/20	Apna
	physiology	Wiring	2/12/20	Ram
	Oral-pathology	Tap Repair	3/12/20	Apna
	prosthodontics	Ac Repair	3/12/20	Rud
	OMFS	lights Repair	2/12/20	Roman
	Orthodontics	Grazer Repair	2/12/20	Roman
<u>9/12/20</u>	public health dentistry	Switch Board - 1	10/12/20	Roman
	Biochemistry	Tap Repair - 1	11/12/20	Apna
	AMR	lights Repair - 1	10/12/20	Rahul
	Cons & Endo	pipe leakage	11/12/20	Rubh
	OMFS	Ac Repair	11/12/20	Rahul

Date	Department	Equipment	Received date	Person
15/12/20	Oral pathology	Projector	16/12/20	Peanecth
	Anatomy	fan Repair	16/12/20	Paman
	Orthodontics	Geyser Repair	17/12/20	Pakel
	Periodontics	Wiring	17/12/20	Pakel
	Physiology	Tap Repair	16/12/20	Kubla
23/12/20	DADH	light Repair - 3	24/12/20	Paman
	prosthodontics	fan Repair	24/12/20	Paman
	pedodontics	pipe Repair	25/12/20	Ola
	OMR	Switch Board	25/12/20	Pakel
	OMPS	fan Repair	24/12/20	Paman
	Cons & Endo	Projector	24/12/20	Paman

January - 2021

date	Department	complaints	Received date	SSN
30/1/21	Anatomy	Cigaretts Repair	5/1/21	Ranjan
	Biochemistry	pipe leakage	4/1/21	Aparna
	Pharmacology	fans Repair	5/1/21	Ranjan
	microbiology	Refrigerator	4/1/21	Bheesh
	Oral pathology	Cigaretts Repair	4/1/21	Ranjan
	PHD	Refrigerator	5/1/21	Bheesh
12/1/21	Periodontics	AC Repair	13/1/21	Bheesh
	Orthodontics	Refrigerator	14/1/21	Bheesh
	Dm	Tap Repair	13/1/21	Aparna
	Physiology	Cigaret Repair	14/1/21	Venkatesh
	OmfS	AC Repair	13/1/21	Venkatesh

1605-160007

Date	Department	Complaint	Rectified date	Sign
21/1/21	OMR	Yaws Repair	22/1/21	Raman
	Orthodontics	lights Repair	27/1/21	Raman
	Pedodontics	de Repair	23/1/21	Ramesh
	PHD	Refrigerator	23/1/21	Ramesh
	Anatomy	lights Repair-3	28/1/21	Raman
	Pharmacology	lights Repair-4	29/1/21	Raman
	General medicine	Yaws Repair-3	29/1/21	Raman
	OMFS	lights Repair-1	29/1/21	Ramesh
	Core & Endo	de Repair	28/1/21	Ramesh
	Orthodontics	de Repair	28/1/21	Ramesh

February

Date	Department	Complaints	Resolved date	Signature
2/2/21	Microbiology	Lights Repair-3	3/2/21	Ramana
	Biochemistry	Pipe leakage	3/2/21	R. Bhagavan
	DADH	Lights Repair	4/2/21	Ramana
	Pathology	Lights Repair	4/2/21	Ramana
	Ophthalmology	Refrigerator	3/2/21	Arul
11/2/21	Pathology	AC Repair	12/2/21	Ramesh
	renal pathology	Lights Repair	12/2/21	Ramana
	Omik	fan Repair	13/2/21	Ramana
	Genecal Pathology	Lights Repair	13/2/21	Ramana
	PHD	Switch Board	13/2/21	Venkat

		Complained	Received date	status
12/2/21	Development	Wiring	18/2/21	visited
	Anatomy	lights Repair	18/2/21	Room
	Biochemistry	pipe leakage	19/2/21	Appo
	microbiology	Air Repair	19/2/21	Brush
	OMF	lights Repair	18/2/21	Room
	com & endo			
23/2/21	Biotechnology	Generator Repair	24/2/21	visited
	General surgery	lights Repair	24/2/21	Room
	Pharmacology	fan Repair	24/2/21	Room
	PHD	spc Repair	25/2/21	Brush
	OMR	Switch Board	25/2/21	Brush

March

Date	Department	Component	Received date	Signature
3/3/21	OMR	Jaws Repair	04/03/21	Ramana
	prosthodontics	Speyers - 1	04/03/21	Balaji
	Microbiology	Switch Board - 1	05/03/21	Venkat
	periodontics	AC Repair - 1	05/03/21	Ramesh
	Oral pathology	Wgts Repair - 2	04/03/21	Ram
	Biochemistry	2-Jaws Repair	04/03/21	Ram
11/3/21	orthodontics	Wgts Repair - 3	12/03/21	Ram
	Endodontics and Conservative	Wiring	12/03/21	Balaji
	public Health dentistry	Switch Board	13/03/21	Balaji
	pedodontics	Projector	13/03/21	Ramath
	Pharmacology	Jaws Repair - 2	12/03/21	Balaji

Date	Department	Complaint	Received date	Status
	Anatomy	Lights Repair - 5	12/03/21	Repair
17/3/21	OMFS	Lights Repair - 2	18/3/21	Repair
	Physiology	fan Repair - 2	18/3/21	Repair
	Gen. pathology	Lights Repair - 1	19/3/21	Repair
	Periodontics	projector - 1	19/3/21	Repair
	Orthodontics	wiring	19/3/21	Repair
27/3/21	Prosthodontics	Switch board	26/3/21	Repair
	Microbiology	fan - 2	26/3/21	Repair
	Oral pathology	fan - 1	26/3/21	Repair
	Biochemistry	Lights Repair - 1	26/3/21	Repair
	Public Health Dentistry	wiring	26/3/21	Repair

August

Date	Department	Complaint	Resolved/Date	Signature
04/08/21	Prosthodontics	Jaw Repair - 3	05/08/21	Panicker
	Orthodontics	Light Repair - 2	05/08/21	Panicker
	Academy	Tap Repair - 2	06/08/21	Panicker
	OMR	Pipe leakage	06/08/21	Panicker
	Cos & Endo	Projector	05/08/21	Panicker
02/08/21	OMR	lights Repair	13/08/21	Panicker
	pedodontics	wiring	13/08/21	Panicker
	oral pathology	Switch Board	14/08/21	Panicker
	perio dentics	AC Repair	14/08/21	Panicker
	Public Health Dentistry	Refrigerator	14/08/21	Panicker

Date	Department	Complaint	Repaired date	Sign
18/08/21	OMIES	lights Repair	19/08/21	Raman
	Orthodontics	Switch Board	19/08/21	Raman
	Physiology	Wiring	19/08/21	Raman
	General Medicine	fans Repair - 2	20/08/21	Raman
	Pharmacology	lights Repair	20/08/21	Raman
	Cons and Endo	Wiring	19/08/21	Raman
24/08/21	Biochemistry	lights Repair	25/08/21	Raman
	Prosthodontics	fans Repair	25/08/21	Raman
	OAR	Wiring	26/08/21	Raman
	Pedodontics	Switch Board	26/08/21	Raman
	Oral pathology	Pipe leakage	25/08/21	Raman
	Public Health Dentistry	Refrigerator	25/08/21	Raman

September

Date	Department	Complaint	Rectified date	Signature
02/09/21	OMFS	Yaws Repair - 3'	03/09/21	Roman
	physiology	Tap Repair	03/09/21	Appasa
	periodontics	Wiring	03/09/21	Roman
	Cous & Endo	Refrigerator	04/09/21	P. Seeth
	prosthodontics	pipe leakage	04/09/21	Appasa
08/09/21	OMR	Tap Repair	09/09/21	Appasa
	Anatomy	pipe leakage	09/09/21	Appasa
	Gen. pathology	Switch Board - 1	10/09/21	Venkat
	Orthodontics	wiring	10/09/21	Venkat
	Pharmacology	light Repair	09/09/21	Roman
	pedodontics	pipe Repair	10/09/21	Roman

Date	Department	Complaints	Resolved date	By
17/09/21	public health dentistry	Cigarettes Repair - 1	18/09/21	Ramona
	pedodontics	glass Repair - 2	18/09/21	Ramona
	Oral pathology	Tap Repair - 1	19/09/21	Affaw
	Microbiology	Refrigerator	19/09/21	Busnel
	prosthodontics	pipe Repair - 2	18/09/21	Wubhu
	OMES	A/c Repair	18/09/21	Busnel
28/09/21	Cous and Endodontics	projector	29/09/21	Blanca
	pedodontics	A/c Repair	29/09/21	Busnel
	orthodontics	projector	30/09/21	Blanca
	Anatomy	Cigaret Repair - 3	29/09/21	Ramona
	OMES	glass Repair - 1	30/09/21	Ramona

Maintenance Checklist - 2020

Procedure.	Aug	Sep	Oct	Nov	Dec
① clean pan strainer.	✓	✓	✓	✓	✓
② clean & flush pan.	✓		✓	✓	
③ Check bleed off valve to make sure it re operative.	✓	✓	✓	✓	✓
④ Lubricate pump and pump motor according to manufacturer's instruction.	✓			✓	
⑤ Check operating level in pan & adjust float valve if necessary.	✓	✓	✓	✓	✓
⑥ Check water distribution system & spray pattern.	✓	✓	✓	✓	✓
⑦ Check drift eliminator.	✓				✓
⑧ Check the fan blades for cracks, missing balancing weight & vibration.	✓				✓

Aug. Sep. Oct. Nov. Dec.

- | | Aug. | Sep. | Oct. | Nov. | Dec. |
|---|------|------|------|------|------|
| ⑨ check sheaves & bushings for corrosion. Scrape & coat with ZRC | ✓ | | | | |
| ⑩ lubricate fan shaft bearings | ✓ | | | ✓ | |
| ⑪ lubricate fan motor bearings | | | ✓ | | |
| ⑫ check belt tension & adjust | ✓ | ✓ | ✓ | ✓ | ✓ |
| ⑬ inspect & slide grease sliding motor base | | | | | ✓ |
| ⑭ check fan screens, inlet louvers, fans & (dry) cooler coil. | ✓ | ✓ | ✓ | ✓ | ✓ |
| ⑮ inspect & clean protective finish.
→ galvanized: scrape & coat with ZRC
→ stainless: clean & polish with stainless cleaner. | | | ✓ | | |
| ⑯ check water quality for biological contamination | ✓ | ✓ | ✓ | ✓ | ✓ |

Maintenance Check-list - 2021

Procedure

	Jan	Feb	March	Apr	May	Jun	July	Aug	Sep
① clean Pan stained	✓	✓	✓	✓	✓	✓	✓	✓	✓
② clean and flush pan	✓			✓	✓		✓		✓
③ check bleed-off valve to make sure it is operative	✓	✓	✓	✓	✓	✓	✓	✓	✓
④ lubricate Pump and pump motor according to manufacturer's instructions	✓				✓			✓	
⑤ check operating level in pan and adjust float valve if necessary	✓	✓	✓	✓	✓	✓	✓	✓	✓
⑥ check operating level in pan	✓	✓	✓	✓	✓	✓	✓	✓	✓
⑥ check water distribution system and spray pattern	✓	✓	✓	✓	✓	✓	✓	✓	✓
⑦ check drift eliminator	✓					✓			✓
⑧ check the fan blades for cracks, missing balancing weights and vibrations	✓		✓						✓

Maintenance Checklist

	Jan	Feb	Mar	Apr	May	Jun	Jul	Aug	Sep	Oct	Nov	Dec
9) check sheaves and bushings for composition. scrape and coat with ZRC												✓
10) lubricate fan shaft bearings			✓									
11) lubricate fan motor bearings												
12) check belt tension and adjust	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓
13) inspect and grease sliding motor base												
14) check fan screens, inlet louvers, fan and (dry) cooler coils	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓
15) inspect and clean protective finish - Galvanized: scrape and coat with ZRC - Stainless: clean and polish with a stainless steel cleaner												
16) check water quality for biological contamination.	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓